



COLLEGE OF VETERINARY AND ANIMAL SCIENCES

(earlier known as the Postgraduate Institute of Veterinary & Animal Sciences)
OPP. DR. PDKV CAMPUS, KRISHI NAGAR, AKOLA 444 104 (MS)
(MAHARASHTRA ANIMAL AND FISHERY SCIENCES UNIVERSITY, NAGPUR)



No. AD/COVAS/UG College/2350/2026

Date: 28/01/2026

NOTICE/ ADVERTISEMENT FOR CALLING QUOTATION

Subject: Invitation of sealed quotations for supply of Biosafety cabinet (Laminar Flow) Class II Type B2 (100% Exhaust) with UV Timer & Hour Meter, Virus Burn Out Unit and Servo controlled Stabilizer to the College of Veterinary & Animal Sciences, Akola.


With respect to subject cited above, sealed quotations are invited from interested suppliers for supply of Biosafety cabinet (Laminar Flow) Class II Type B2 (100% Exhaust) with UV Timer & Hour Meter, Virus Burn Out Unit and Servo controlled Stabilizer to the College of Veterinary & Animal Sciences, Akola with following specifications.

SN	Name of Item	Specifications	Quantity
1	Biosafety cabinet (Laminar Flow) Class II Type B2 (100% Exhaust) with UV Timer & Hour Meter, Virus Burn Out Unit and Servo controlled Stabilizer with warranty	<ol style="list-style-type: none"> Material of construction – Stainless Steel Grade 304 with 1.5 mm thickness. Work Area in ft. - 4' x 2' x 2' Internal Dimension: L x D x H (mm) : 1235 x 600 x 650 +/- 5 mm Down Flow Velocity: 60 FPM at 4" from the ULPA Grill Inflow Flow Velocity: 100 FPM across the work access opening. Body Finish: Glossy Design: Class II – B2 compliant 100% Air exhausted back to the environment through a dedicated duct. Cleanliness Class: ISO Standard 14644 -1 ISO Class 3 Noise Level: Less than 60 Db at 1 Meter from the face of the filter Stages of filtration: <ol style="list-style-type: none"> ULPA Final = 0.12μ with an efficiency of 99.9995% ULPA Exhaust = 0.12μ with an efficiency of 99.9995% Primary Pre Filter = 15 micron with efficiency of 85% Front Door: One piece full view frame less sliding toughened glass door 6 mm thick. Motorized automatic door. ULPA Filter Grill: SS 304 grills duly perforated and polished. Motor Blower Assembly for Supply and Exhaust: variable speed motor mounted on PU coated FRP Blowers with Aluminum Impellers. Switches & Speed Controller: Feather touch microprocessor control. Fluorescent lighting: 36 Watt Slim tube lights Ultra Violet Lighting: 3 feet length G30T8 30 Watt Electrical Socket: 5 / 15 Amps socket – 1/2 Nos. Internal Work Area: Complete Stainless Steel 304 with rounded joint less corners. Pressure gauge: Magnehelic gauge Gas / Vacuum Cock Wheels & Leveling Lugs: PU solid wheels with 	2

	<ul style="list-style-type: none"> plated brackets & Metal Leveling Lugs. 22. Virus Burn Out Unit 23. Suitable Servo controlled Stabilizer 24. Power Supply & Service: Single phase 230V, 50Hz power supply 25. All services from Front and Back <ul style="list-style-type: none"> • Alarm for Excess Height opening and automatic cutoff for UV tube when the door is open. • Interlocked switches for UV & Lights. • Stainless Steel drain collection system. • Raised arm rest prevents grille blocking, and provides comfortable working posture. • Single-piece work tray to contain spillage with sloped easy-to-swipe perimeter. • System is ergonomically designed with angled cabinet front which is sloped at around 10° for enhanced comfort and reduced operator fatigue. • For easy servicing all cabinet components, including HEPA filters are easily accessible from the front to allow for rapid service and minimal work disruption during preventative maintenance and recertification. • System shall comply with NSF/ANSI 49 to ensure a safe operating environment for Class II, Type B2 conditions. 	
--	---	--

Terms and Conditions:

1. Outward no. and date of this notice should be written as reference on the quotation as well as on the envelope.
2. '**Quotation for supply of Biosafety cabinet (Laminar Flow) Class II Type B2 to the College of Veterinary & Animal Sciences, Akola**' should be superscripted on the sealed envelope.
3. The rates should be inclusive of all taxes and FOR at the Department of Veterinary Microbiology/Public Health, COVAS, Akola and shall be valid for next six months.
4. The bidder has to submit **EMD/Bank Guaranty for Rs. 95000/-** in the name of **the Associate Dean, COVAS, Akola** along with the quotation. As per Government Resolution dated 01/12/2016, MSME are exempted on submission of valid document.
5. Successful bidder needs to submit **3 % Security Deposit** for one year in the name of **the Associate Dean, COVAS, Akola**.
6. Deviation in the specification may be liable for rejection and the decision of the Associate Dean, COVAS, Akola will be final.
7. The supplier should be responsible for supply of the goods/material in good and working conditions within fifteen days period from the issuance of supply order, installation and training.
8. All rights regarding accepting or rejecting the quotation(s) and Purchase Order for part/complete supply are reserved with the Associate Dean, COVAS, Akola and no reason will be assigned for such act.
9. The supplier should enclose supplementary/supporting documents such as OEM/ valid OEM Authorization, three Yrs of Experience, ITR for last three Yrs/ Annual turnover, PAN card, GST Certificate, etc. with quotation.
10. The supplier should enclose Undertaking (on Letter Head) in the name of The Associate Dean, COVAS, Akola. The format of the same is enclosed.
11. The quotes should be submitted in the name of the **Associate Dean, COVAS, Akola**, in sealed envelope **on or before 05/02/2026 (17:00 hrs)**. Late and incomplete quotations will be rejected.


 Associate Dean
 COVAS, Akola

Copy to:

- 1) Notice Board, COVAS, Akola / Department of Veterinary Microbiology, COVAS, Akola.
- 2) In-charge, Intranet Cell, COVAS, Akola with request to upload on COVAS website for wider publicity.

(On Letter Head of the Bidder)

Undertaking

I/We, the undersigned, M/s. _____ residing at/ located at _____, hereby provide this undertaking that I/we am/are interested in providing the goods/ materials/ services mentioned in Advertisement No. _____ published by your office.

I/We have submitted only one quotation/ tender for the said advertisement. Furthermore, I/we do not have any personal or vested interests with the office publishing the advertisement or with any person working within that office.

Signature and Stamp of the Firm